



*Pride in Service with Integrity*

BROWARD SHERIFF'S OFFICE  
DEPARTMENT OF FIRE RESCUE  
AND EMERGENCY SERVICES



# SAFETY WARNING

## AMSINO IV DRIPS

SW 08-02

**SAFETY  
WARNING**

Fire Marshal's Bureau  
2601 West Broward Boulevard  
Fort Lauderdale, FL 33312


Phone: 954-831-8210

Fax: 954-831-8218

E-mail: [FMO\\_Issues@sheriff.org](mailto:FMO_Issues@sheriff.org)

**Date: April 30, 2008**

**Logistics received a report from a municipal partner that some AMSINO IV Drips are marked and packaged incorrectly.**



It has been reported that AMSINO IV Drip packages marked 60gtts actually contain 10 drop sets.

All units should check their inventory of Amsino IV Drips both 60gtts and 10 gtts for LOT # 121507A.

If any packages with this LOT # are located they should be further examined to insure that the proper drip set is contained in each package. All incorrectly marked or packaged IV drip sets should be packaged, marked with the station in which they were located and returned to Logistics.